

7. X-RAY DETECTORS

images at different exposure times for each position of the crystal, extensible software would make it easier to set up the experiment. Finally, the software should permit access to all of the readout modes of the detector. For instance, a detector may be capable of rapidly scanning a small region of interest for alignment purposes, or it may be capable of streak-mode operation for certain types of time-resolved experiments. Available CCD detector software for macromolecular applications has room for much improvement. Hopefully, software will continue to undergo rapid development. Standardization is especially needed.

7.2.5. Applications to macromolecular crystallography

Storage rings. CCD detectors have gained widespread acceptance for macromolecular crystallography at storage-ring sources, in part because of the high-quality data they give, but more for their speed, convenience and efficiency. Accurate data to high resolution are especially important for MAD phasing, and CCD detectors excel in this application. In the past with film, or even storage phosphors, teams of perhaps ten people were required to perform a synchrotron experiment; today, a single person per shift can perform an experiment. With increasing beam flux, improved X-ray optics and faster CCDs, it is often possible to collect full data sets in little more than an hour. Anticipated improvements in speed for CCD detectors should soon make it feasible to collect fine-sliced rotation data routinely; these data are expected to yield better structure solutions.

Home laboratories. Acceptance of CCD detectors for macromolecular crystallography at home laboratories has been slower, in part because there is not such a premium on speed, and in part because of cost. Diffracted spot sizes are larger than at synchrotrons, so highly accurate data should be obtainable. Fully automatic storage phosphor systems work quite well with conventional sources and at this time are lower in cost than large CCD detectors. However, they have a minimum cycle time, caused by the mechanics of the readout scheme, and the required exposure for a strongly diffracting crystal can best this time by a wide margin. Thus, for strongly diffracting specimens, CCD detectors can be significantly more efficient.

7.2.6. Future of CCD detectors

The basic principles of CCD detector technology are now well developed, but various incremental improvements have already been demonstrated and may be expected in commercial detectors. These include larger detector areas, faster read times (owing to both faster electronics and multi-amplifier CCDs), more flexible control electronics, better optimized phosphors and calibrations, and, especially, better software. Lower-cost CCD detectors would certainly be welcome. It is easily predicted that the application of CCD detectors will continue to increase rapidly for at least several more years until displaced by even better technologies, such as pixel array detectors (see Section 7.1.4 in Chapter 7.1).

References

7.1

- Amemiya, Y., Matsushita, T., Nakagawa, A., Satow, Y., Miyahara, J. & Chikawa, J.-I. (1988). *Design and performance of an imaging plate system for X-ray diffraction study.* *Nucl. Instrum. Methods Phys. Res. A*, **266**, 645–653.
- Arndt, U. W. (1991). *Second-generation X-ray television area detectors.* *Nucl. Instrum. Methods Phys. Res. A*, **310**, 395–397.
- Arndt, U. W., Gilmore, D. J. & Wonacott, A. J. (1977). *X-ray film.* In *The rotation method in crystallography*, edited by U. W. Arndt & A. J. Wonacott, pp. 207–218. Amsterdam: North-Holland Publishing Co.
- Barbosa, A. F., Gabriel, A. & Craievich, A. (1989). *An X-ray gas position-sensitive detector – construction and characterization.* *Rev. Sci. Instrum.* **60**, 2315–2317.
- Barna, S. L., Shepherd, J. A., Tate, M. W., Wixted, R. L., Eikenberry, E. F. & Gruner, S. M. (1997). *Characterization of prototype pixel array detector (PAD) for use in microsecond framing time-resolved X-ray diffraction studies.* *IEEE Trans. Nucl. Sci.* **44**, 950–956.
- Barna, S. L., Tate, M. W., Gruner, S. M. & Eikenberry, E. F. (1999). *Calibration procedures for charge-coupled device X-ray detectors.* *Rev. Sci. Instrum.* **70**, 2927–2934.
- Blum, M., Metcalf, P., Harrison, S. C. & Wiley, D. C. (1987). *A system for collection and on-line integration of X-ray diffraction data from a multiwire area detector.* *J. Appl. Cryst.* **20**, 235–242.
- Charpak, G. (1982). *Parallax-free, high-accuracy gaseous detectors for X-ray and VUV localization.* *Nucl. Instrum. Methods*, **201**, 181–192.
- Datte, P., Beuville, E., Millaud, J. & Xuong, N.-H. (1999). *A digital pixel address generator for pixel array detectors.* *Nucl. Instrum. Methods Phys. Res. A*, **421**, 492–501.
- Eikenberry, E. F., Tate, M. W., Bilderback, D. H. & Gruner, S. M. (1992). *X-ray detectors: comparison of film, storage phosphors and CCD detectors.* *Inst. Phys. Conf. Ser.* **121**, 273–280.
- Farrell, R., Vanderpuye, K., Cirignano, L., Squillante, M. R. & Entine, G. (1994). *Radiation detection performance of very high-gain avalanche photodiodes.* *Nucl. Instrum. Methods Phys. Res. A*, **353**, 176–179.
- Fujita, H., Tsai, D.-Y., Itoh, T., Doi, K., Morishita, J., Ueda, K. & Ohtsuka, A. (1992). *A simple method for determining the modulation transfer-function in digital radiography.* *IEEE Trans. Med. Imaging*, **11**, 34–39.
- Gramsch, E., Szawłowski, M., Zhang, S. & Madden, M. (1994). *Fast, high-density avalanche photodiode-array.* *IEEE Trans. Nucl. Sci.* **41**, 762–766.
- Gruner, S. M., Milch, J. R. & Reynolds, G. T. (1978). *Evaluation of area photon detectors by a method based on detective quantum efficiency (DQE).* *IEEE Trans. Nucl. Sci.* **NS-25**, 562–565.
- Hall, G. (1995). *Silicon pixel detectors for X-ray diffraction studies at synchrotron sources.* *Q. Rev. Biophys.* **28**, 1–32.
- Hamlin, R., Cork, C., Howard, A., Nielsen, C., Vernon, W., Matthews, D. & Xuong, N. H. (1981). *Characteristics of a flat multiwire area detector for protein crystallography.* *J. Appl. Cryst.* **14**, 85–93.
- Iles, G., Raymond, M., Hall, G., Lovell, M., Seller, P. & Sharp, P. (1996). *Hybrid pixel detector for time resolved X-ray diffraction experiments at synchrotron sources.* *Nucl. Instrum. Methods Phys. Res. A*, **381**, 103–111.
- Krause, K. L. & Phillips, G. N. Jr (1992). *Experience with commercial area detectors: a 'buyer's' perspective.* *J. Appl. Cryst.* **25**, 146–154.
- Ludewigt, B., Jaklevic, J., Kipnis, I., Rossington, C. & Spieler, H. (1994). *A high-rate, low-noise, X-ray silicon strip detector system.* *IEEE Trans. Nucl. Sci.* **41**, 1037–1041.
- Milch, J. R., Gruner, S. M. & Reynolds, G. T. (1982). *Area detectors capable of recording X-ray diffraction patterns at high count rates.* *Nucl. Instrum. Methods*, **201**, 43–52.
- Moy, J.-P. (1999). *Large area X-ray detectors based on amorphous silicon detector.* *Thin Solid Films*, **337**, 213.
- Rehak, P., Walton, J., Gatti, E., Longoni, A., Sanpietro, M., Kemmer, J., Dietl, H., Holl, P., Klanner, R., Lutz, G., Wylie, A. & Becker, H. (1986). *Progress in semiconductor drift detectors.* *Nucl. Instrum. Methods Phys. Res. B*, **248**, 367–378.

REFERENCES

7.1 (cont.)

- Rossi, G., Renzi, M., Eikenberry, E. F., Tate, M. W., Bilderback, D., Fontes, E., Wixted, R., Barna, S. & Gruner, S. M. (1999). *Tests of a prototype pixel array detector for microsecond time-resolved X-ray diffraction*. *J. Synchrotron Rad.* **6**, 1096–1105.
- Sarvestani, A., Besch, H. J., Junk, M., Meissner, W., Pavel, N., Sauer, N., Stiehler, R., Walenta, A. H. & Menk, R. H. (1998). *Gas amplifying hole structures with resistive position encoding: a new concept for a high rate imaging pixel detector*. *Nucl. Instrum. Methods Phys. Res. A*, **419**, 444–451.

7.2

- Barna, S. L., Tate, M. W., Gruner, S. M. & Eikenberry, E. F. (1999). *Calibration procedures for charge-coupled device X-ray detectors*. *Rev. Sci. Instrum.* **70**, 2927–2934.
- Coleman, C. I. (1985). *Imaging characteristics of rigid coherent fiber optic tapers*. *Adv. Electron. Electron Phys.* **64B**, 649–661.
- Deckman, H. W. & Gruner, S. M. (1986). *Format alterations in CCD based electro-optic X-ray detectors*. *Nucl. Instrum. Methods Phys. Res. A*, **246**, 527–533.
- Eikenberry, E. F., Tate, M. W., Belmonte, A. L., Lowrance, J. L., Bilderback, D. & Gruner, S. M. (1991). *A direct-coupled detector for synchrotron X-radiation using a large format CCD imaging array*. *IEEE Trans. Nucl. Sci.* **38**, 110–118.
- Gruner, S. M., Barna, S. L., Wall, M. E., Tate, M. W. & Eikenberry, E. F. (1993). *Characterization of polycrystalline phosphors for area X-ray detectors*. *Proc. SPIE*, **2009**, 98–108.
- Hammersley, A. P., Brown, K., Burmeister, W., Claustre, L., Gonzalez, A., McSweeney, S., Mitchell, E., Moy, J.-P., Svensson, S. O. & Thompson, A. W. (1997). *Calibration and application of an X-ray image intensifier/charge-coupled device detector for*

- monochromatic macromolecular crystallography*. *J. Synchrotron Rad.* **4**, 67–77.
- Moy, J.-P. (1994). *A 200 mm input field 5–80 keV detector based on an X-ray image intensifier and CCD camera*. *Nucl. Instrum. Methods Phys. Res. A*, **348**, 641–644.
- Moy, J.-P. (1998). *Image quality of scintillator based X-ray electronic imagers*. *Proc. SPIE*, **3336**, 187–194.
- Moy, J. P., Hammersley, A. P., Svensson, S. O., Thompson, A., Brown, K., Claustre, L., Gonzalez, A. & McSweeney, S. (1996). *A novel technique for accurate intensity calibration of area X-ray detectors at almost arbitrary energy*. *J. Synchrotron Rad.* **3**, 1–5.
- Naday, I., Ross, S., Kanyo, M., Westbrook, M., Westbrook, E. M., Phillips, W. C., Stanton, M. J. & O'Mara, D. (1995). *The gold detector: modular CCD area detector for macromolecular crystallography*. *Proc. SPIE*, **2415**, 236–249.
- Shepherd, J. A., Gruner, S. M., Tate, M. W. & Tecotzky, M. (1995). *A study of persistence in gadolinium oxysulfide X-ray phosphors*. *Proc. SPIE*, **2519**, 24–30.
- Stevens, A. & Schrama-de Pauw, A. (1974a). *Vapour-deposited CsI:Na layers. I. Morphologic and crystallographic properties*. *Philips Res. Rep.* **29**, 340–352.
- Stevens, A. & Schrama-de Pauw, A. (1974b). *Vapour-deposited CsI:Na layers. II. Screens for application in X-ray imaging devices*. *Philips Res. Rep.* **29**, 353–362.
- Tate, M. W., Eikenberry, E. F., Barna, S. L., Wall, M. E., Lowrance, J. L. & Gruner, S. M. (1995). *A large-format high-resolution area X-ray detector based on a fiber-optically bonded charge-coupled device (CCD)*. *J. Appl. Cryst.* **28**, 196–205.
- Tate, M. W., Eikenberry, E. F. & Gruner, S. M. (1997). *Coupling format variations in X-ray detectors based on charge-coupled devices*. *Rev. Sci. Instrum.* **68**, 47–54.