

10.1. INTRODUCTION TO CRYOCRYSTALLOGRAPHY

Both Si diodes and Pt resistance sensors have become sufficiently miniaturized to make them preferred choices at the lowest temperatures. Thermocouples are acceptable above about 80 K. A reliable method of calibration makes use of the known temperature of a phase transition of a crystal in the normal data-collection position. For example, KH_2PO_4 (often referred to as KDP) has a sharp transition at 123 K from tetragonal to orthorhombic, and is commonly used. Another possibility is KH_2AsO_4 , which has a corresponding phase transition at 95 K.

Two readout temperatures suffice, one at room temperature and one at the phase transition. The difference between readout temperature and crystal-site temperature can be assumed to vary linearly with T , so interpolation or extrapolation is simple.

10.1.4.4. Transfer of the crystal to the diffractometer

Inspection of Figs. 10.1.4.1–10.1.4.4 reveals that the mounting of a crystal on a mounting pin *via* the traditional placement of the pin in the hole of a standard goniometer head is not simple, because the cooling nozzle is in the way. The solution to the problem is a design that allows side entry. This is most commonly achieved with a magnetic platform on the goniometer head and a corresponding magnetic base on the mounting pin, but an alternative means of side entry employs a slot on a modified goniometer head; the slot is equipped with a spring-loaded catch that allows a very smooth, but stable, catch of the pin.

The use of liquid- N_2 cooling and side entry, and the requirement of reproducible knowledge of crystal temperature at all times, led to the development of a set of tools for crystal mounting as described by Parkin & Hope (1998). The tools include special transfer tongs used for moving crystals from liquid N_2 to the goniometer head. The temperature of the crystal is maintained by the heat capacity and low heat conductance of the tongs. The operation is independent of the orientation of the goniometer head because there is no liquid to contain.

10.1.4.5. Automated robotic crystal handling

The era of structural genomics has necessitated a move to high throughput at high-intensity synchrotron sources. To meet this goal, a number of robotic crystal-handling devices have been developed. These in turn have had the beneficial effect of increased standardization in mounting-pin geometry.

10.1.5. Concluding note

With correctly functioning low-temperature equipment and appropriate techniques, a crystal can be maintained frost-free for the duration of a data-collection run. Formation of frost on the crystal indicates malfunction of the equipment, or operator error. The most likely cause is operator error, but faulty equipment cannot be ruled out. The techniques described here have been used for collecting thousands of data sets from ice-free crystals and crystal mounts. There is no reason to accept frost problems as an unavoidable part of cryocrystallography.

References

- Bald, W. B. (1984). *The relative efficiency of cryogenic fluids used in the rapid quench cooling of biological samples*. *J. Microsc.* **134**, 261–270.
- Bellamy, H. D., Phizackerley, R. P., Soltis, S. M. & Hope, H. (1994). *An open-flow cryogenic cooler for single-crystal diffraction experiments*. *J. Appl. Cryst.* **27**, 967–970.
- Brinkmann, C., Weiss, M. S. & Weckert, E. (2006). *The structure of the hexagonal crystal form of hen egg-white lysozyme*. *Acta Cryst.* **D62**, 349–355.
- Chinte, U., Shah, B., Chen, Y.-S., Pinkerton, A. A., Schall, C. A. & Hanson, B. L. (2007). *Cryogenic (<20 K) helium cooling mitigates radiation damage to protein crystals*. *Acta Cryst.* **D63**, 486–492.
- Corbett, M. C., Latimer, M. J., Poulos, T. L., Sevrioukova, I. F., Hodgson, K. O. & Hedman, B. (2007). *Photoreduction of the active site of the metalloprotein putidaredoxin by synchrotron radiation*. *Acta Cryst.* **D63**, 951–960.
- Douzou, P., Hui Bon Hoa, G. & Petsko, G. A. (1975). *Protein crystallography at sub-zero temperatures: lysozyme–substrate complexes in cooled mixed solutions*. *J. Mol. Biol.* **96**, 367–380.
- Esnouf, R. M., Ren, J., Garman, E. F., Somers, D. O'N., Ross, C. K., Jones, E. Y., Stammers, D. K. & Stuart, D. I. (1998). *Continuous and discontinuous changes in the unit cell of HIV-1 reverse transcriptase crystals on dehydration*. *Acta Cryst.* **D54**, 938–953.
- Fu, Z.-Q., Du Bois, G. C., Song, S. P., Harrison, R. W. & Weber, I. T. (1999). *Improving the diffraction quality of MTCP-1 crystals by post-crystallization soaking*. *Acta Cryst.* **D55**, 5–7.
- Garman, E. F. & Mitchell, E. P. (1996). *Glycerol concentrations required for cryoprotection of 50 typical protein crystallization solutions*. *J. Appl. Cryst.* **29**, 584–587.
- Garman, E. F. & Nave, C. (2009). *Synchrotron radiation damage in protein crystals examined under various conditions by different methods*. *J. Synchrotron Rad.* **16**, 129–132.
- Götz, G., Mészáros, E. & Vali, G. (1991). *Atmospheric particles and nuclei*, p. 142. Budapest: Akadémiai Kiadó.
- Haas, D. J. & Rossmann, M. G. (1970). *Crystallographic studies on lactate dehydrogenase at -75°C* . *Acta Cryst.* **B26**, 998–1004.
- Harp, J. M., Timm, D. E. & Bunick, G. J. (1998). *Macromolecular crystal annealing: overcoming increased mosaicity associated with cryocrystallography*. *Acta Cryst.* **D54**, 622–628.
- Holton, J. M. (2009). *A beginner's guide to radiation damage*. *J. Synchrotron Rad.* **16**, 133–142.
- Hope, H. (1988). *Cryocrystallography of biological macromolecules: a generally applicable method*. *Acta Cryst.* **B44**, 22–26.
- Hope, H. (1990). *Cryocrystallography of biological macromolecules at ultra-low temperature*. *Annu. Rev. Biophys. Biophys. Chem.* **19**, 107–126.
- Hope, H., Frolow, F., von Böhlen, K., Makowski, I., Kratky, C., Halfon, Y., Danz, H., Webster, P., Bartels, K. S., Wittmann, H. G. & Yonath, A. (1989). *Cryocrystallography of ribosomal particles*. *Acta Cryst.* **B45**, 190–199.
- Hui Bon Hoa, G. & Douzou, P. (1973). *Ionic strength and protein activity of supercooled solutions used in experiments with enzyme systems*. *J. Biol. Chem.* **248**, 4649–4654.
- Juers, D. H. & Matthews, B. W. (2004). *The role of solvent transport in cryo-annealing of macromolecular crystals*. *Acta Cryst.* **D60**, 412–421.
- Kriminski, S., Caylor, C. L., Nonato, M. C., Finkelstein, K. D. & Thorne, R. E. (2002). *Flash-cooling and annealing of protein crystals*. *Acta Cryst.* **D58**, 459–471.
- Leidenfrost, J. G. (1756). *De aquae communis nonnullis qualitatibus tractatus*. Duisburg, Germany.
- Meents, A., Wagner, A., Schneider, R., Pradervand, C., Pohl, E. & Schulze-Briese, C. (2007). *Reduction of X-ray-induced radiation damage of macromolecular crystals by data collection at 15 K: a systematic study*. *Acta Cryst.* **D63**, 302–309.
- Mitchell, E. P. & Garman, E. F. (1994). *Flash freezing of protein crystals: investigation of mosaic spread and diffraction limit with variation of cryoprotectant concentration*. *J. Appl. Cryst.* **27**, 1070–1074.
- Parkin, S. & Hope, H. (1998). *Macromolecular cryocrystallography: cooling, mounting, storage and transportation of crystals*. *J. Appl. Cryst.* **31**, 945–953.
- Parkin, S. & Hope, H. (2003). *Low-temperature water reconstruction in concanavalin A, with implications for controlled protein crystal annealing*. *Acta Cryst.* **D59**, 2228–2236.
- Petrova, T., Ginell, S., Mitschler, A., Hazemann, I., Schneider, T., Cousido, A., Lunin, V. Y., Joachimiak, A. & Podjarny, A. (2006). *Ultrahigh-resolution study of protein atomic displacement parameters at cryotemperatures obtained with a helium cryostat*. *Acta Cryst.* **D62**, 1535–1544.
- Petsko, G. A. (1975). *Protein crystallography at sub-zero temperatures: cryoprotective mother liquors for protein crystals*. *J. Mol. Biol.* **96**, 381–392.
- Post, B., Schwartz, R. S. & Fankuchen, I. (1951). *An improved device for X-ray diffraction studies at low temperatures*. *Rev. Sci. Instrum.* **22**, 218–220.